

SUBJECT Physical & Microbiological Test

TEST LOCATION TÜV SÜD China


TÜV SÜD Products Testing (Shanghai) Co., Ltd. B-3/4, No.1999 Du Hui Road, Minhang District

Shanghai 201108, P.R. China

CLIENT NAME Shandong Zhushi Pharmaceutical Group Co., Ltd.

CLIENT ADDRESS No.6 Shande Road, Shan County, Heze City, Shandong, China

TEST PERIOD 15-Apr-2020~29-Apr-2020



Note: (1) General Terms & Conditions as mentioned overleaf. (2) The results relate only to the items tested.(3) The test report shall not be reproduced except in full without the written approval of the laboratory.(4) Without the agreement of the laboratory, the client is not authorized to use the test results for unapproved propaganda.

P.R. China



TEST REPORT

Sample Description : Disposable Surgical Face Mask

Sample Quantity : 50 pieces
Lot Number/Batch Code : Lot20041002

Specification :

Size : 17.5cm×9.5cm

Brand Name : /

Remark: The above information was provided by applicant.

Summary of Test Results

No.	Test Item	Test Standard	Test Standard Type II R	Judgement
1	Bacterial Filtration Efficiency Test (BFE), %	EN 14683:2019+AC:2019(E) Annex B	≥ 98	Pass
2	Differential Pressure Test (Pa/cm²)	EN 14683:2019+AC:2019(E) Annex C	< 60	Pass
3	Synthetic Blood Penetration Test (kPa)	ISO 22609:2004	≥ 16.0	Pass
4	Microbial Cleanliness Test (CFU/g)	EN 14683:2019+AC:2019(E) Annex D	≤ 30	Pass

Note: Pass = Meet customer requirements;

Fail = Fail customer requirements;

= No comment; N.D. = Not detected.

Photo of Samples



Chemical/Microbiology Laboratory: TÜV SÜD Products Testing (Shanghai) Co., Ltd.

B-3/4, No.1999 Du Hui Road, Minhang District Shanghai 201108 P.R. China Phone: +86 (21) 6037 6375 Fax: +86 (21) 6037 6345 Email: food.chem@tuv-sud.cn Webpage: www.tuv-sud.cn Regional Head Office: TÜV SÜD Certification and Testing (China) Co., Ltd. No.151 Heng Tong Road Shanghai 200 070 _ P.R.China



Results

No.	Test Item	Test Result	
		Specimen 1#: 99.6%	
1		Specimen 2#: 99.6%	
	Bacterial Filtration Efficiency (BFE) Test	Specimen 3#: 99.6%	
		Specimen 4#: 99.6%	
		Specimen 5#: 99.6%	
2	Differential Pressure Test	34.2 Pa/cm ²	
3	Synthetic Blood Penetration Test	Specimen 1#~32#: None seen	
4		Specimen 1#: <1 CFU/g	
		Specimen 2#: 2 CFU/g	
	Microbial Cleanliness Test	Specimen 3#: <1 CFU/g	
		Specimen 4#: <1 CFU/g	
		Specimen 5#: 2 CFU/g	

Bacterial Filtration Efficiency (BFE) Test

1. Purpose

For evaluating the bacterial filtration efficiency (BFE) of mask.

2. Sample description was given by client

Sample description : Disposable Surgical Face Mask

Specification :

Lot Number : Lot20041002

Sample Receiving Date: 2020-04-15

3. Test Method

EN 14683:2019+AC:2019(E) Annex B

4. Apparatus and materials

- 4.1 Staphylococcus aureus ATCC 6538.
- 4.2 Peptone water.
- 4.3 Tryptic Soy Broth(TSB).
- 4.4 Tryptic Soy Agar(TSA).
- 4.5 Bacterial filtration efficiency test apparatus.
- 4.6 Six-stage viable particle Anderson sampler.
- 4.7 Flow meters.

5. Test specimen

- 5.1 As requested by client, take a total of 5 test specimens.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4 h at (21±5)°C and (85±5)% relative humidity.

Chemical/Microbiology Laboratory: TÜV SÜD Products Testing (Shanghai) Co., Ltd.

B-3/4, No.1999 Du Hui Road, Minhang District Shanghai 201108

P.R. China

Phone: +86 (21) 6037 6375 Fax: +86 (21) 6037 6345 Email: food.chem@tuv-sud.cn Webpage: www.tuv-sud.cn Regional Head Office: TÜV SÜD Certification and Testing (China) Co., Ltd. No.151 Heng Tong Road Shanghai 200 070 _ P.R.China







6. Procedure

- 6.1 Preparation of the bacterial challenge: Dilute the cultutre in peptone water to achieve a concentration of approximately 5×10⁵ CFU/mL.
- 6.2 Adjust the flow rate through the Anderson sampler to 28.3 L/min.
- 6.3 Deliver the challenge to the nebulizer using a syringe pump. Purge tubing and nebulizer of air bubbles.
- 6.4 Perform a positive control run without a test specime to determine the number of viable aerosol particles being generated. The mean particle size (MPS) of the aerosol will also be calculated from the results of these positive control plates.
 - 6.4.1 Initiate the aerosol challenge by turning on the air pressure and pump connected to the nebulizer. Immediaterly begin sampling the aerosol using the Anderson sampler.
 - 6.4.2 Time the challenge suspension to be delivered to the nebulizer for 1 min.
 - 6.4.3 Time the air pressure and Anderson sampler to run for 2 min.
 - 6.4.4 At the conclusion of the positive control ran, remove plates from the Anderson sampler.
- 6.5 Place new agar plates into Anderson sampler and clamp the test specimen into the top of the Anderson sampler, with the inside of the specimen facing towards the bacterial challenge (test area: 77cm²).
- 6.6 Repeat the challenge procedure for each test specimen.
- 6.7 Repeat a positive control after completion of the sample set.
- 6.8 Perform a negative control run by collecting a 2 min sample of air from the aerosol chamber. No bacterial challenge should be pumped into the nebulizer during the collection of the negative control.
- 6.9 Incubate agar plates at (37±2)°C for (20 to 52) h.
- 6.10 Count each of the six-stage plates of the Anderson sampler.

7. Calculation

Total the count from each of the six plates for the test specimens and positive controls, as specified by the manufacture of Anderson sampler. The filtration efficiency percentages are calculated as follows:

 $BFE=(C-T)/C \times 100$

T is the total plate count for the test specimen.

C is the mean of the total plate counts for the two positive controls.

Phone: +86 (21) 6037 6375

Fax: +86 (21) 6037 6345 Email: food.chem@tuv-sud.cn

Webpage: www.tuv-sud.cn




8. Test results*

P Value Stage Number	Positive Control (A)	Positive Control (B)	Negative Control	Specimen 1#	Specimen 2#	Specimen 3#	Specimen 4#	Specimen 5#
1	23	56	0	0	0	0	0	0
2	52	150	0	0	0	0	0	0
3	134	198	0	0	0	0	0	0
4	230	308	0	0	0	0	0	0
5	1241	1314	0	7	6	5	6	5
6	670	606	0	2	5	5	5	4
Total (T), CFU	2350	2632	<1	9	11	10	11	9
Average (C), CFU	2.5 x10 ³ =	(P _A +P _B) / 2						
BFE ,%			99.6	99.6	99.6	99.6	99.6	
Requirements		//	1	≥	98			1
Remarks	P is the value of corresponding corrected particle counts as specified by the manufacturer of the cascade impactor. T is the total of P value for the test specimen. C is the mean of the total of P value of the two positive controls.							



Phone: +86 (21) 6037 6375 Fax: +86 (21) 6037 6345 Email: food.chem@tuv-sud.cn

Webpage: www.tuv-sud.cn



Differential pressure Test

1.Purpose

The purpose of the test was to measure the differential pressure of masks.

2. Sample description was given by client

Sample description : Disposable Surgical Face Mask

Specification :

Lot Number : Lot20041002 Sample Receiving Date : 2020-04-15

3.Test Method

EN 14683:2019+AC:2019(E) Annex C

4. Apparatus and materials

Differential pressure testing instrument

5.Test specimen

- 5.1 Test specimen are complete masks or shall be cut from masks. Each specimen shall be able to provide 5 different circular test areas of 2.5 cm in diameter.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4 h at (21±5) °C and (85±5)% relative humidity.

6. Procedure

- 6.1 Without a specimen in place, the holder is closed and the differential manometer is zeroed. The pump is started and the flow of air adjusted to 8 L/min.
- 6.2 The pretreated specimen is placed across the orifice (total area 4.9cm², test area diameter 25mm) and clamped into place so as to minimize air leaks.
- 6.3 Due to the presence of an alignment system the tested area of the specimen should be perfectly in line and across the flow of air.
- 6.4 The differential pressure is read directly.
- 6.5 The procedure described in steps 6.1-6.4 is carried out on 5 different areas of the mask and readings averaged.

Results:

P.R. China

Specimen	Test Results* (Pa/cm²)	Average (Pa/cm²)	Requirements	Judgement
1#	33.2			
2#	34.5			
3#	32.1	34.2	< 60	Pass
4#	34.9			
5#	36.3			

Chemical/Microbiology Laboratory: TÜV SÜD Products Testing (Shanghai) Co., Ltd.

B-3/4, No.1999 Du Hui Road, Minhang District Shanghai 201108 Phone: +86 (21) 6037 6375 Fax: +86 (21) 6037 6345 Email: food.chem@tuv-sud.cn Webpage: www.tuv-sud.cn Regional Head Office: TÜV SÜD Certification and Testing (China) Co., Ltd. No.151 Heng Tong Road Shanghai 200 070 _ P.R.China



Synthetic Blood Penetration Test

1.Purpose

For evaluation of resistance of masks to penetration by a fixed volume of synthetic blood at a high velocity.

2. Sample description was given by client

Sample description : Disposable Surgical Face Mask

Specification : /

Lot Number : Lot20041002 Sample Receiving Date : 2020-04-15

3.Test Method

ISO 22609:2004

4. Apparatus and materials

- 4.1 Synthetic blood.
- 4.2 Tensiometer.
- 4.3 Synthetic blood penetration test apparatus;
- 4.4 Targeting plate.
- 4.5 Air pressure source.
- 4.6 Ruler.
- 4.7 Balance.
- 4.8 Controlled temperature and humidity chamber.

5.Test specimen

- 5.1 As requested by client, take a total of 32 test specimens.
- 5.2 Prior to testing, condition all test specimens for a minimum of 4h at (21±5)°C and (85±5) % relative humidity.

6.Procedure

Shanghai 201108

P.R. China

- 6.1 Prepare the synthetic blood (40~44 mN/m) for the test.
- 6.2 Determine the density of the synthetic blood.
- 6.3 Fill the reservoir with new synthetic blood.
- 6.4 Position the test specimen 30.5 cm (12 in.) from the exit of the canula.
- 6.5 Set the reservoir pressure to the approximate pressure.
- 6.6 Place the targeting plate approximately 1 cm away from the mask.
- 6.7 Set the valve timer to 0.5 s. Collect and weigh the amount of fluid delivered (before the targeting

Chemical/Microbiology Laboratory: TÜV SÜD Products Testing (Shanghai) Co., Ltd. B-3/4, No.1999 Du Hui Road, Minhang District Phone: +86 (21) 6037 6375 Fax: +86 (21) 6037 6345 Email: food.chem@tuv-sud.cn Webpage: www.tuv-sud.cn Regional Head Office: TÜV SÜD Certification and Testing (China) Co., Ltd. No.151 Heng Tong Road Shanghai 200 070 P.R.China

の人をきれり章

hole).

- 6.8 Set the valve timer to 1.5 s. Collect and weigh the amount of fluid delivered (before the targeting hole).
- 6.9 Calculate the difference in weight of the two spurts. For a test fluid with a density of 1.003, Table 1 gives the target difference in weight plus lower and upper limits for a velocity range within 2% of the target.

Table 1 Target weight difference

Fluid Pressure (mmHg)	Weight difference for 1s difference in spurt duration (g)			
	Min.	Target	Max.	
120	3.002	3.063	3.124	

- 6.10 Adjust the reservoir pressure and repeat steps 6.7 to 6.9 until the weight difference is within the target range.
- 6.11 Record the weight difference for the spurts exiting the nozzle.
- 6.12 Record the pressure in the reservoir.
- 6.13 Set the valve time to 0.5 s. Collect and weigh the amount of fluid passing through the targeting hole.
- 6.14 Set the valve time to 1.5 s. Collect and weigh the amount of fluid passing through the targeting hole.
- 6.15 The difference in weight between the 0.5 s and 1.5 s spurts through the targeting plate shall be within $+2\% \sim -5\%$ of the difference in weight from the nozzle.
- 6.16 If the differential weight is less than 95 % of the weight difference exiting the nozzle, check the aim of the stream to make sure it is passing cleanly through the targeting hole.
- 6.17 If the differential weight is more than 102 % of the weight difference exiting the nozzle, repeat the weight measurements exiting the nozzle (steps 6.7 to 6.11).
- 6.18 For standard synthetic blood, the timer duration can be estimated using the formula: (p is the density of the test fluid.) $t = 0.5 + (2 \times p g \text{ at } 0.5 \text{ s}) / (g \text{ at } 1.5 \text{ s} g \text{ at } 0.5 \text{ s})$.
- 6.19 Record the timer setting to use as the starting point for subsequent testing.
- 6.20 Mount a test specimen on the specimen holding fixture. If the mask contains pleats, spread the pleats out when mounting the mask onto the fixture to present a single layer of material as the target area.
- 6.21 Squirt the synthetic blood onto the test specimen for the calculated time. Ensure that the synthetic blood hits the target area of mask.
- 6.22 Inspect the inside surface for synthetic blood penetration within 10 s of squirting the synthetic blood against the target area.
- 6.23 Report the results (none / penetration) for each test specimen at the test pressure.

Phone: +86 (21) 6037 6375

+86 (21) 6037 6345

Shanghai

P.R. China

201108



Results:

suits.			
Specimen	Test Results*	Requirements	Judgement
1#	None Seen		Pass
2#	None Seen		Pass
3#	None Seen		Pass
4#	None Seen		Pass
5#	None Seen		Pass
6#	None Seen		Pass
7#	None Seen		Pass
8#	None Seen		Pass
9#	None Seen]	Pass
10#	None Seen		Pass
11#	None Seen		Pass
12#	None Seen		Pass
13#	None Seen		Pass
14#	None Seen	Pass Pressure at 16.0 kPa	Pass
15#	None Seen		Pass
16#	None Seen		Pass
17#	None Seen	(120mmHg)	Pass
18#	None Seen		Pass
19#	None Seen	DEID /	Pass
20#	None Seen	5UD /	Pass
21#	None Seen		Pass
22#	None Seen		Pass
23#	None Seen		Pass
24#	None Seen		Pass
25#	None Seen]	Pass
26#	None Seen]	Pass
27#	None Seen]	Pass
28#	None Seen]	Pass
29#	None Seen]	Pass
30#	None Seen	1	Pass
31#	None Seen	1	Pass
32#	None Seen	1	Pass

Chemical/Microbiology Laboratory: TÜV SÜD Products Testing (Shanghai) Co.,

B-3/4, No.1999 Du Hui Road, Minhang District Shanghai 201108 P.R. China

Phone: +86 (21) 6037 6375 Fax: +86 (21) 6037 6345 Email: food.chem@tuv-sud.cn Webpage: www.tuv-sud.cn

Regional Head Office: TÜV SÜD Certification and Testing (China) Co., Ltd. No.151 Heng Tong Road Shanghai 200 070 P.R.China





TÜV

Microbial Cleanliness Test

1. Purpose

The purpose of the test was to measure microbial cleanliness of mask.

2. Sample description was given by client

Sample description : Disposable Surgical Face Mask

Specification : /

Lot Number : Lot20041002 Sample Receiving Date : 2020-04-15

3. Test Method

According to EN ISO 11737-1:2018 to determine the microbial cleanliness of mask material, and refer to the procedure as described in EN 14683:2019+AC:2019(E) Annex D

4. Apparatus and materials

- 4.1 Orbital shaker.
- 4.2 0.45 um filter.
- 4.3 Tryptic Soy Agar (TSA).
- 4.4 Sabouraud Dextrose Ager (SDA) with chloramphenicol.
- 4.5 Formula of Extraction Liquid: 1g/L peptone, 5g/L NaCl and 2g/L Tween 20.
- 4.6 Extraction apparatus.

5. Test specimen

- 5.1 As requested by client, take a total of 5 mask samples.
- 5.2 Mask samples for testing are provided in the original primary packaging.
- 5.3 Condition at (18 to 26) °C and (45 to 65)% relative humidity during testing.

6. Procedure

- 6.1 Five test specimens are selected from the top, bottom and 3 randomly chosen marks.
- 6.2 The mask is aseptically removed from the packaging and placed in a sterile 500 mL bottle containing 300 mL of extraction liquid.
- 6.3 The bottle is laid down on an orbital shaker and shaken for 5 min at 250 rpm.
- 6.4 After extracting, 100mL of the extraction liquid is filtered through a 0.45 um filter and laid down on a TSA plate for the total viable aerobic microbial count. Another 100 mL aliquot of the same extraction liquid is filtered in the same way and the filter plated on SDA for fungi enumeration.
- 6.5 The plates are incubated for 3 days at 30°C and 7 days at (20 to 25)°C for TSA and SDA plates respectively.
- 6.6 Calculate the colonies of each agar plate.

7. Calculation

For each test specimen calculate the microbial cleanliness as follows by counting the total colonies of the TSA and SDA plates.

201108

P.R. China



Results*:

Specimen	Colonies of the TSA Plate	Colonies of the SDA Plate	Microbial Cleanliness, (CFU/g)	Requirements	Judgement
1#	0	0	<1		
2#	0	2	2	EN 14683:2019+AC:2019(E) Annex D	
3#	0	0	<1		Pass
4#	0	0	<1	EN ISO 11737-1:2018 ≤ 30 CFU/g	
5#	0	2	2		

Note:

1.*denotes this test was carried out by external laboratory assessed as competent.

Phone: +86 (21) 6037 6375

Fax: +86 (21) 6037 6345

Webpage: www.tuv-sud.cn

Email: food.chem@tuv-sud.cn

2.This report is for internal use only such as internal scientific research ,education, quality control, product R&D.



